


SPS Secondary Emission Monitors



ADC: Full Scale = 10v
Sem counts (FS) = 2047

$$E = \frac{\text{Number of secondary electrons emitted by detector foil}}{\text{Number of protons traversing detector foil}}$$

$$V = \frac{N_p * E * Q_{el} * G}{C} \quad \frac{N_p}{\text{count}} = \frac{V_{(FS)} * C}{Q_{el} * E * S * G} \quad \text{With } E=3.9\% \quad N_p/\text{count} = 6.58 * 10^9$$

- N_p = Number of protons traversing detector foil
- E = Secondary emission efficiency
- C = Capacitance of integrator (33nF)
- Q_{el} = Electron charge $1.602189 * 10^{-19}$ Coul.
- V = Output of amplifier (V)
- S = Sem counts (FS) = 2047 counts
- G = Amplifier gain = 3.92

SEM electronics:

Gain error between channels +/- 0.5%
Noise cables plus detector +/- 1.5 count

Detector foils:

Diameter 145mm
Aluminium foil thickness 20 um
Titanium foil thickness 20 um

Specifications of BCTs installed in TT60

LF cut off 1 Hz
HF cut off 20KHz
Precision +/- 1.5%
Resolution $1.0 * E11p$
Calibrator $1.0 * E13p$